

# Certificate

## EC-Type-Examination

in accordance with Directive 97/23/EC

Certificate No.: 01 202 642TR-B-16 0005

Name and address of the manufacturer:

**CIRCOR Instrumentation  
dba Hoke Incorporated  
405 Centura Ct.  
USA - 29303 Spartanburg, SC**

It is herewith certified that the EC-Type Example mentioned below meets the requirements of the Directive 97/23/EC.

Tested acc. to Directive 97/23/EC: **EC-Type-Examination (Module B)**

Test report No.: 642-124835888-10

Description of Type Example:

### Safety Valve

Design: direct acting spring loaded safety valve  
Type: 6711L4YA (10 – 12 bar)  
6711L4YB (17 – 21 bar)  
6711L4YC (24 – 28 bar)  
6711L4YD (37 – 41 bar)  
Medium: for analysis gases service, free from dust (fluid groups 1 and 2)  
Size: ¼"-NPT  
Pressure: 10 - 41 barg (150 - 600 psi)


Production Site:

CIRCOR Instrumentation  
dba Hoke Incorporated  
405 Centura Ct.  
USA - 29303 Spartanburg, SC

Valid until: 03. June 2025

The CE marking must not be affixed and the Declaration of Conformity not be issued prior to completion of the corresponding conformity assessment procedure according to Directive 97/23/EC.

Aachen, 18. March 2016

  
Dipl. Ing. Manfred Rönning

TÜV Rheinland-Certification Body  
for Pressure Equipment  
TÜV Rheinland Industrie Service GmbH  
Notified Body, ID-No. 0035  
Am Grauen Stein, D-51105 Köln

Außenstelle  
Krefelder Str. 225, 52070 Aachen, Tel. ++49-241/1825-870  
E-Mail: service-is-mw@de.tuv.com



642-124835888-10-zert 16-02-29 Circor-Hoke\_6711

# Certificate

## EC-Type-Examination

in accordance with Directive 97/23/EC

Certificate No.: 01 202 642-B-14-0069

Name and address of the manufacturer:

**Circor Instrumentation Technologies**  
**dba Hoke, Inc.**  
405 Centura Court  
Spartanburg, SC 29303 USA

It is herewith certified that the EC-Type Example mentioned below meets the requirements of the Directive 97/23/EC.

Tested acc. to Directive 97/23/EC: **EC-Type-Examination (Module B)**

Test report No.: 642-14-0069, 2nd supplement

Description of Type Example:

**Safety Valve**

Design: direct acting spring loaded safety valve  
Type: LR6000, MR6000, HR6000, XR6000  
Size: 1/4"-NPT, 1/4", 6/8/12 mm Gyrolock  
Pressure: 0.5 – 413 barg (7 – 6000 psig)

Production Site:

**Circor Instrumentation Technologies**  
**dba Hoke, Inc.**  
405 Centura Court  
Spartanburg, SC 29303 USA

Valid until: 14<sup>th</sup> December 2024

The CE marking must not be affixed and the Declaration of Conformity not be issued prior to completion of the corresponding conformity assessment procedure according to Directive 97/23/EC.

Aachen, 06. March 2015

  
Dipl. Ing. Manfred Rönig

TÜV Rheinland-Certification Body  
for Pressure Equipment  
TÜV Rheinland Industrie Service GmbH  
Notified Body, ID-No. 0035  
Am Grauen Stein, D-51105 Köln

Außenstelle  
Krefelder Str. 225, 52070 Aachen, Tel. ++49-241/1825-870  
E-Mail: service-is-mw@de.tuv.com



SIV HOKE\_USA 14-0069 EG-B-Cert.docx

www.tuv.com

 **TÜVRheinland®**  
Genau. Richtig.